



Operation Instructions for Electronic Manifold Meters



HT-750

Table of Contents

Summary	1
Cautions	1
Product Structures	2
Button Functions	4
Initial Operation	6
An Introduction to Display	7
Description of Settings	9
Description of Symbols	12
Routine Measurement	13
Leakage Test	13
Product Specification	14

Summary



This product is the integration of functions of pressure measurement and temperature measurement. It is mainly used for leakage detection, refrigerant charging, troubleshooting and maintenance of the refrigeration system.

Cautions

To ensure the proper use of this product, please read the operation instructions carefully:

1. Never disassemble or refit this product personally without authorization.
2. If the hose is damaged, please change it in a timely manner.
3. Do not drop or impact the instrument to avoid damage.
4. The measuring instrument can only be used by qualified and approved personnel.
5. If this product lies idle for long, please take out the battery to avoid battery leakage and corrosion.
6. Refrigerant gas will do hazard to the environment, so it is required to ensure that it will comply with the local environmental regulations when using it.
7. If there is any failure in the instrument, stop using it and contact the supplier or manufacturer in time.
8. Do not discard the waste batteries at will, but send them to the specified collection points for waste batteries.
9. Please evaluate whether the measuring object or environment is safe, and note the safety standards in your measuring areas.
10. The instrument is suitable for the measurement of most non corrosive refrigerants, water and alcohol media. The instrument may not be used in explosion-proof areas.


Product structures



- ① Terminal T1 temperature clamping port
- ② Terminal T2 temperature clamping port
- ③ Display screen
- ④ Button
- ⑤ Liquid view mirror (used to observe the flow of refrigerants)

- ① Terminal T1 temperature clamping port
- ② Terminal T2 temperature clamping port
- ③ Display screen
- ④ Button
- ⑤ Liquid view mirror (used to observe the flow of refrigerants)
- ⑥ High pressure end control valve: loosen it counterclockwise, turn on the value for ventilation, tighten it clockwise, and there is no air ventilation after the valve is turned off.
- ⑦ Liquid filling pipe support
- ⑧ High pressure end connector
- ⑨ Liquid filling end connector: used to connect the refrigerant bottles
- ⑩ Low pressure end connector
- ⑪ Loosen it counterclockwise, turn on the value for ventilation, tighten it clockwise, and there is no air ventilation after the valve is turned off.
- ⑫ Battery compartment



Button functions



Buttons	Functions
	Go to or exit the leakage test interface
	Clear pressure readings
	Power on/off Backlight on/off
	Go to the refrigerant selection interface; Start/Stop leakage test
	Up button: to switch interfaces; Change parameters during setup
	Down button: to switch interfaces; Change parameters during setup
	Go to the setting interface
	View maximum/minimum/ mean values
	Return to the measurement interface

Initial operation



Power on/off

Long press the “ /  ” button for power on or off.

Clear pressure

Before connecting the hose, clear the pressure readings: Turn on the value, and press “P=0” button for clearance.

Backlight





Shortly press the “ /  ” button to activate backlight, and press it once again to deactivate backlight.

Mounting the battery





Open the battery cover. Install 4 AA alkaline batteries. Note the polarity.

They may not be mounted wrongly. After battery installation, cover the battery cover.

Maximum, minimum and mean values


Press the “ ” button to display the maximum value → continue to press the “ ” button to display the minimum value → continue to press the “ ” button to display the mean value → continue to press the “ ” button to exit (or press “ESC” button to quit).

Select refrigerants


Press the “ ” button to go to the refrigerant selection interface. Press the “ ” or the “ ” button to switch refrigerants. Press the “ ” button again to return to the measurement interface.


An Introduction to display interface


If no refrigerant has been chosen, the following interface is displayed by default when powered on




If one of refrigerants is selected, the following interface is displayed by default when powered on



Press the “ ” button to switch to the over cooling and overheating display interface, as shown in the figure below:




Continue to press the “▼” button to switch to the evaporating temperature and condensing temperature display interface.



Pressure is displayed at the upper part of the screen, and the evaporating temperature and condensing temperature are displayed at the lower part of the screen.

Continue to press the “▼” button to switch to the temperature difference display interface.



$$\Delta t = T_2 - T_1$$

Description of settings

Continuously press the “SET” button 6 times, with the display in turn: pressure unit setting interface → temperature unit setting interface → relative/absolute pressure setting interface → Air pressure setting interface → refrigerating/heating mode setting interface → automatic power off function setting interface. Finally, press the “ESC” button to quit the setting interface.


1. Setting pressure units

From the pressure unit setting interface, press the “▲” button or “▼” button to switch the pressure units. There are four pressure units in all: kpa, Mpa, psi and bar




2. Setting temperature units

From the temperature unit setting interface, press the “▲” button or “▼” button to switch temperature units. There are 2 temperature units in all: °C and °F.



3. Setting relative/absolute pressure

From the relative/absolute pressure setting interface, press the “▲” button or “▼” button to select the relative pressure measurement mode or absolute pressure measurement mode.



Absolute pressure: pressure value by adding air pressure.

Relative pressure: pressure value minus air pressure.


If the relative pressure measurement mode is chosen: according to the selected pressure unit, prel or psig is displayed at the left upper corner of the screen.

If the absolute measurement is chosen: according to the selected pressure measurement unit, pabs or psia is displayed at the left upper corner of the screen.

4. Air pressure setting interface


There is difference in air pressure in different areas. The user will make adjustments to air pressure at the place where he or she is currently located.

From the air pressure setting interface, press the “▲” button or “▼” button to adjust the air pressure value.



5. Setting the refrigerating/heating modes

From the refrigerating/heating mode setting interface, press the “▲” button or “▼” button to select one of the modes. There are three modes in all: 🔥 hot pump mode, ❄️ refrigerating mode, and 🔥❄️ automatic mode.






6. Setting automatic power off functions

From the automatic power off function setting interface, press “▲” button or “▼” button, select “on” to turn on the automatic power off function, and select “off” to turn off the automatic power off function. When the automatic power off function is turned on: if you do not press the button within 20 minutes, it will be automatically shut down.



Description of Symbols

Symbols	Meaning
T1	Number of temperature probe
T2	Number of temperature probe
prel or psig	Relative pressure
pabs or psia	Absolute pressure
	Hot pump mode
	Refrigerating mode
	Auto mode
max	Maximum value
min	Minimum value
mean	Mean value
p=0	Clear pressure readings
Δt	T2-T1
Δt_{oh}	Overheat
SH	Evaporating pressure
Δt_{cu}	Over cooling
SC	Condensing pressure
t_o	Refrigerant's evaporating temperature
E_v	Refrigerant's evaporating pressure
t_c	Refrigerant's condensing temperature
C_o	Refrigerant's condensing pressure
t_{oh}	Actually measured temperature at T1 terminal (evaporating temperature)
t_{cu}	Actually measured temperature at T2 terminal (condensing temperature)

Routine measurement


1. Connect 2 temperature clamps respectively to the temperature probe connector of the instrument.
2. Power on
3. Select refrigerants: press the "**R**" button to enter the refrigerant selection interface. Press the "**▲**" button or "**▼**" button to switch refrigerants. Press the "**R**" button once again to return to the measurement interface.
4. Select the measurement mode (please refer to "Description of Settings" for the operation method)
5. Connect the high pressure end and the low pressure end under an ambient pressure.
6. If the reading is not zero without pressure input, please press the "P = 0" button to clear the pressure readings.
7. Turn off the control valve before connecting the refrigerant hose: Connect the
8. Connect the liquid filling pipe to the system.
9. Input pressure to the measuring instrument.
10. Read the readings. Press the "**▲**" button or "**▼**" button to be able to switch to other measurement interfaces.

Leakage test


This function is used for leakage test in refrigeration or hot pump system. Leakage analysis is performed by system pressure over a period of time.

1. Press the "P = 0" button to clear the pressure readings (clearance if there is no pressure input).
2. Use a hose to connect the high pressure end of the instrument to the system.
3. Press the "mode" button to switch to the leakage test interface, as shown in the figure below:

4. Press the "▶ R ■" button to start the test, and press the "▶ R ■" once again to stop the test.



5. The test duration depends on the system tested.



Product specification

Pressure measuring range (relative pressure: -100 to 6000Kpa/-0.1 to 6Mpa/-14.5 to 870psi / -1 to 60bar)

Overload pressure: 6500Kpa/6.5Mpa/940psi/65bar

Pressure accuracy: $\pm 0.5\%$ FS

Pressure resolution: 1Kpa, 0.01bar, 0.1psi, 0.001Mpa (0.01Mpa during negative pressure)

Temperature measuring range: -10-200°C/14-392°F

Temperature accuracy: $\pm (1^\circ\text{C}+1\text{digit})$;

Temperature resolution: 0.1°C/0.1°F

Optional refrigerants (83 kinds): R11, R12, R123, R124, R125, R1270, R13, R134A, R14, R141b, R142b, R143A, R152A, R170, R21, R218, R22, R227E, R23, R236E, R245C, R245, R290, R32, R401A, R401b, R401C, R402A, R402b, R403A, R403b, R404A, R405A, R406A, R407A, R407b, R407C, R407d, R407E, R408A, R409A, R409b, R41, R410A, R410b, R11A, R411b, R412A, R413A, R414A, R414b, R415A, R415b, R416A, R417A, R418A, R419A, R420A, R421A, R421b, R422A, R422b, R422C, R422d, R423A, R424A, R425A, R426A, R427A, R428A, R50, R500, R501, R502, R503, R504, R507A, R508A, R508b, R509A, R600, R600A, R717

Operating temperature: 0 to 45°C

Storage temperature: -20 to 60°C

Ambient humidity: 10%RH to 90%RH

Power supply: 4 AA batteries

Dimension: 204mm×116mm×65mm

Accessories: tool box, hose (each in red, yellow and blue), Operation

Instructions, 4 AA alkaline batteries, temperature clamps x2